

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043568.D
 Acq On : 28 Oct 2024 17:14
 Operator : JC/MD
 Sample : PB164394TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164394TB

Quant Time: Oct 29 03:26:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	157154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	298510	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	123854	49.426	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.860%
35) Dibromofluoromethane	5.385	113	97712	48.272	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	351010	50.094	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	11.079	95	125219	48.227	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	26778	26.910	ug/l	97
20) Methylene Chloride	2.788	84	4022	1.875	ug/l #	96
43) Isopropyl Acetate	6.342	43	113795	23.614	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043568.D
Acq On : 28 Oct 2024 17:14
Operator : JC/MD
Sample : PB164394TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164394TB

Quant Time: Oct 29 03:26:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

