

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043569.D
 Acq On : 28 Oct 2024 17:38
 Operator : JC/MD
 Sample : PB164394ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164394ZHE#01

Quant Time: Oct 29 03:27:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	213325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	397289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165192	48.564	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.120%
35) Dibromofluoromethane	5.385	113	126690	47.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.060%
50) Toluene-d8	8.647	98	465524	49.919	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.840%
62) 4-Bromofluorobenzene	11.079	95	168510	48.764	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.520%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	44965	33.289	ug/l	96
20) Methylene Chloride	2.782	84	3674	1.262	ug/l	91
43) Isopropyl Acetate	6.342	43	149910	23.374	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043569.D
Acq On : 28 Oct 2024 17:38
Operator : JC/MD
Sample : PB164394ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164394ZHE#01

Quant Time: Oct 29 03:27:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

