

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043570.D
 Acq On : 28 Oct 2024 18:01
 Operator : JC/MD
 Sample : PB164394ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164394ZHE#02

Quant Time: Oct 29 03:27:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	211615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	393848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	345503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161991	48.008	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.020%
35) Dibromofluoromethane	5.385	113	125163	46.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	8.647	98	458033	49.545	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.080%
62) 4-Bromofluorobenzene	11.079	95	168261	49.117	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.240%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	44921	33.525	ug/l	97
20) Methylene Chloride	2.782	84	3429	1.187	ug/l #	78
43) Isopropyl Acetate	6.342	43	148542	23.363	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043570.D
Acq On : 28 Oct 2024 18:01
Operator : JC/MD
Sample : PB164394ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164394ZHE#02

Quant Time: Oct 29 03:27:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

