

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043571.D
 Acq On : 28 Oct 2024 18:24
 Operator : JC/MD
 Sample : PB164394ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164394ZHE#03

Quant Time: Oct 29 03:28:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	225140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	427240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	369828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	171560	47.790	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.580%		
35) Dibromofluoromethane	5.385	113	134884	46.558	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.120%		
50) Toluene-d8	8.647	98	502872	50.143	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.280%		
62) 4-Bromofluorobenzene	11.079	95	178633	48.069	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.140%		
Target Compounds					Qvalue	
16) Acetone	2.392	43	47258	33.150	ug/l	98
20) Methylene Chloride	2.782	84	3804	1.238	ug/l	98
43) Isopropyl Acetate	6.342	43	151271	21.932	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043571.D
Acq On : 28 Oct 2024 18:24
Operator : JC/MD
Sample : PB164394ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164394ZHE#03

Quant Time: Oct 29 03:28:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

