

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043580.D
 Acq On : 28 Oct 2024 21:52
 Operator : JC/MD
 Sample : PB164394ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164394ZHE#12

Quant Time: Oct 29 03:32:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	234786	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	437918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	178109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165988	44.338	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.680%
35) Dibromofluoromethane	5.385	113	133793	45.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	8.647	98	506396	49.264	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.079	95	185247	48.634	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.260%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	51613	34.718	ug/l	99
18) Methyl Acetate	2.715	43	4507	1.037	ug/l	98
20) Methylene Chloride	2.788	84	4074	1.272	ug/l #	87
43) Isopropyl Acetate	6.343	43	156153	22.088	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043580.D
Acq On : 28 Oct 2024 21:52
Operator : JC/MD
Sample : PB164394ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164394ZHE#12

Quant Time: Oct 29 03:32:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

