

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102921\
 Data File : VX025017.D
 Acq On : 29 Oct 2021 11:27
 Operator : JC/MD
 Sample : VX1029WBL05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK623

Quant Time: Oct 30 01:08:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 30 01:06:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	167307	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.055	117	150323	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	65381	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	43490	40.779	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.560%
7) Chloroethane-d5	1.666	69	33385	45.796	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.306	63	85348	32.476	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.960%
21) 2-Butanone-d5	4.464	46	103878	97.514	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.510%
24) Chloroform-d	5.068	84	127096	45.545	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.080%
26) 1,2-Dichloroethane-d4	5.958	65	94659	46.257	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.520%
32) Benzene-d6	5.982	84	201156	47.145	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.300%
36) 1,2-Dichloropropane-d6	7.312	67	69240	47.618	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.240%
41) Toluene-d8	8.653	98	188422	46.159	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.320%
43) trans-1,3-Dichloroprop...	8.951	79	38605	46.199	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.400%
47) 2-Hexanone-d5	9.390	63	73416	99.168	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.170%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	95237	48.436	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.880%
66) 1,2-Dichlorobenzene-d4	12.323	152	63593	50.125	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102921\
Data File : VX025017.D
Acq On : 29 Oct 2021 11:27
Operator : JC/MD
Sample : VX1029WBL05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK623

Quant Time: Oct 30 01:08:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 30 01:06:24 2021
Response via : Initial Calibration

