

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043584.D
 Acq On : 29 Oct 2024 10:06
 Operator : JC/MD
 Sample : VX1029MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029MBL01

Quant Time: Oct 30 01:25:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	157445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	292323	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	261965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	125443	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	109724	43.706	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.420%
35) Dibromofluoromethane	5.379	113	91513	46.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.340%
50) Toluene-d8	8.647	98	347478	50.640	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	11.079	95	129743	51.027	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043584.D
Acq On : 29 Oct 2024 10:06
Operator : JC/MD
Sample : VX1029MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1029MBL01

Quant Time: Oct 30 01:25:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

