

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043591.D
 Acq On : 29 Oct 2024 12:56
 Operator : JC/MD
 Sample : P4578-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-10252024

Quant Time: Oct 30 01:30:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	156843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	288755	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	110702	44.265	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.520%
35) Dibromofluoromethane	5.385	113	88640	45.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.540%
50) Toluene-d8	8.647	98	337480	49.790	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.580%
62) 4-Bromofluorobenzene	11.079	95	114524	45.598	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.200%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1722	1.734	ug/l	95
20) Methylene Chloride	2.788	84	642	0.300	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043591.D
Acq On : 29 Oct 2024 12:56
Operator : JC/MD
Sample : P4578-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-10252024

Quant Time: Oct 30 01:30:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

