

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043592.D
 Acq On : 29 Oct 2024 13:19
 Operator : JC/MD
 Sample : P4548-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP BLANK

Quant Time: Oct 30 01:30:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	160173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	298248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	113453	44.422	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.840%
35) Dibromofluoromethane	5.385	113	92267	45.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.240%
50) Toluene-d8	8.647	98	348344	49.758	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	123224	47.500	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.000%
Target Compounds						
					Qvalue	
3) Chloromethane	1.288	50	583	0.266	ug/l #	86
16) Acetone	2.392	43	1574	1.552	ug/l #	82
20) Methylene Chloride	2.788	84	780	0.357	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043592.D
Acq On : 29 Oct 2024 13:19
Operator : JC/MD
Sample : P4548-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP BLANK

Quant Time: Oct 30 01:30:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

