

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043594.D
 Acq On : 29 Oct 2024 14:06
 Operator : JC/MD
 Sample : P4548-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Oct 30 01:31:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	149761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	275286	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	102170	42.785	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.580%		
35) Dibromofluoromethane	5.385	113	85635	45.875	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.740%		
50) Toluene-d8	8.647	98	320898	49.660	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.320%		
62) 4-Bromofluorobenzene	11.079	95	114189	47.689	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.380%		
Target Compounds						Qvalue
4) Vinyl Chloride	1.374	62	1134	0.597	ug/l	99
8) Diethyl Ether	2.136	74	607	0.566	ug/l	77
16) Acetone	2.392	43	1991	2.100	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	1940	0.896	ug/l	89
65) Chlorobenzene	10.079	112	4463	0.863	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043594.D
Acq On : 29 Oct 2024 14:06
Operator : JC/MD
Sample : P4548-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Oct 30 01:31:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

