

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043604.D
 Acq On : 29 Oct 2024 17:57
 Operator : JC/MD
 Sample : P4600-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190-GW-498-500

Quant Time: Oct 30 01:36:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	237971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89594	43.817	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.640%
35) Dibromofluoromethane	5.385	113	74107	45.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.840%
50) Toluene-d8	8.647	98	279025	49.951	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	11.079	95	105956	51.189	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.380%
Target Compounds						
16) Acetone	2.392	43	8608	10.601	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043604.D
Acq On : 29 Oct 2024 17:57
Operator : JC/MD
Sample : P4600-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190-GW-498-500

Quant Time: Oct 30 01:36:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

