

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102924\
 Data File : VX043608.D
 Acq On : 29 Oct 2024 19:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 30 01:38:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.535	0.487	9.0	94	0.00
3 P	Chloromethane	0.684	0.630	7.9	94	0.00
4 C	Vinyl Chloride	0.634	0.622	1.9#	96	0.00
5 T	Bromomethane	0.259	0.257	0.8	96	0.00
6 T	Chloroethane	0.223	0.220	1.3	95	0.00
7 T	Trichlorofluoromethane	0.775	0.787	-1.5	99	0.01
8 T	Diethyl Ether	0.358	0.379	-5.9	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.524	0.521	0.6	97	0.00
10 T	Methyl Iodide	0.757	0.781	-3.2	97	0.00
11 T	Tert butyl alcohol	0.149	0.157	-5.4	98	-0.04
12 CM	1,1-Dichloroethene	0.536	0.539	-0.6#	98	0.00
13 T	Acrolein	0.116	0.115	0.9	103	0.00
14 T	Allyl chloride	0.926	0.917	1.0	94	0.00
15 T	Acrylonitrile	0.336	0.353	-5.1	99	0.00
16 T	Acetone	0.317	0.294	7.3	92	0.00
17 T	Carbon Disulfide	1.120	1.013	9.6	87	0.00
18 T	Methyl Acetate	0.925	0.905	2.2	95	0.00
19 T	Methyl tert-butyl Ether	2.037	2.113	-3.7	99	0.00
20 T	Methylene Chloride	0.682	0.675	1.0	99	0.00
21 T	trans-1,2-Dichloroethene	0.566	0.578	-2.1	100	0.00
22 T	Diisopropyl ether	1.915	1.951	-1.9	98	0.00
23 T	Vinyl Acetate	1.582	1.677	-6.0	97	0.00
24 P	1,1-Dichloroethane	1.101	1.100	0.1	97	0.00
25 T	2-Butanone	0.462	0.462	0.0	93	0.00
26 T	2,2-Dichloropropane	0.901	0.791	12.2	87	0.00
27 T	cis-1,2-Dichloroethene	0.723	0.766	-5.9	101	0.00
28 T	Bromochloromethane	0.524	0.520	0.8	100	0.00
29 T	Tetrahydrofuran	0.300	0.301	-0.3	95	0.00
30 C	Chloroform	1.154	1.146	0.7#	96	0.00
31 T	Cyclohexane	0.842	0.834	1.0	97	0.00
32 T	1,1,1-Trichloroethane	0.969	0.953	1.7	94	0.00
33 S	1,2-Dichloroethane-d4	0.797	0.760	4.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.339	0.345	-1.8	101	0.00
36 T	1,1-Dichloropropene	0.402	0.389	3.2	98	0.00
37 T	Ethyl Acetate	0.488	0.496	-1.6	96	0.00
38 T	Carbon Tetrachloride	0.444	0.421	5.2	93	0.00
39 T	Methylcyclohexane	0.499	0.498	0.2	96	0.00
40 TM	Benzene	1.341	1.365	-1.8	99	0.00
41 T	Methacrylonitrile	0.256	0.271	-5.9	98	0.00
42 TM	1,2-Dichloroethane	0.482	0.481	0.2	95	0.00
43 T	Isopropyl Acetate	0.807	0.813	-0.7	95	0.00
44 TM	Trichloroethene	0.333	0.333	0.0	97	0.00
45 C	1,2-Dichloropropane	0.330	0.345	-4.5#	100	0.00
46 T	Dibromomethane	0.254	0.262	-3.1	99	0.00
47 T	Bromodichloromethane	0.446	0.461	-3.4	95	0.00
48 T	Methyl methacrylate	0.383	0.395	-3.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043608.D
 Acq On : 29 Oct 2024 19:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 30 01:38:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	101	0.00
50 S	Toluene-d8	1.174	1.203	-2.5	100	0.00
51 T	4-Methyl-2-Pentanone	0.492	0.510	-3.7	95	0.00
52 CM	Toluene	0.813	0.855	-5.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.479	0.492	-2.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.542	-6.3	98	0.00
55 T	1,1,2-Trichloroethane	0.339	0.364	-7.4	100	0.00
56 T	Ethyl methacrylate	0.550	0.588	-6.9	98	0.00
57 T	1,3-Dichloropropane	0.563	0.604	-7.3	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.281	-5.6	103	0.00
59 T	2-Hexanone	0.374	0.380	-1.6	93	0.00
60 T	Dibromochloromethane	0.335	0.362	-8.1	97	0.00
61 T	1,2-Dibromoethane	0.351	0.374	-6.6	100	0.00
62 S	4-Bromofluorobenzene	0.435	0.447	-2.8	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.314	0.314	0.0	101	0.00
65 PM	Chlorobenzene	1.076	1.088	-1.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.351	0.367	-4.6	98	0.00
67 C	Ethyl Benzene	1.754	1.752	0.1#	98	0.00
68 T	m/p-Xylenes	0.671	0.684	-1.9	98	0.00
69 T	o-Xylene	0.674	0.696	-3.3	100	0.00
70 T	Styrene	1.120	1.189	-6.2	100	0.00
71 P	Bromoform	0.253	0.259	-2.4	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.542	3.503	1.1	98	0.00
74 T	N-amyl acetate	1.725	1.692	1.9	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.274	1.276	-0.2	98	0.00
76 T	1,2,3-Trichloropropane	1.069	1.032	3.5	96	0.00
77 T	Bromobenzene	0.915	0.947	-3.5	102	0.00
78 T	n-propylbenzene	3.862	3.866	-0.1	96	0.00
79 T	2-Chlorotoluene	2.561	2.518	1.7	99	0.00
80 T	1,3,5-Trimethylbenzene	2.953	2.965	-0.4	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.389	0.381	2.1	91	0.00
82 T	4-Chlorotoluene	2.874	2.873	0.0	99	0.00
83 T	tert-Butylbenzene	3.003	2.980	0.8	97	0.00
84 T	1,2,4-Trimethylbenzene	3.010	3.085	-2.5	99	0.00
85 T	sec-Butylbenzene	3.520	3.565	-1.3	97	0.00
86 T	p-Isopropyltoluene	2.992	3.010	-0.6	96	0.00
87 T	1,3-Dichlorobenzene	1.639	1.673	-2.1	101	0.00
88 T	1,4-Dichlorobenzene	1.699	1.682	1.0	101	0.00
89 T	n-Butylbenzene	2.438	2.504	-2.7	95	0.00
90 T	Hexachloroethane	0.486	0.489	-0.6	94	0.00
91 T	1,2-Dichlorobenzene	1.676	1.725	-2.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.260	0.248	4.6	88	0.00
93 T	1,2,4-Trichlorobenzene	1.017	1.088	-7.0	100	0.00
94 T	Hexachlorobutadiene	0.375	0.362	3.5	96	0.00
95 T	Naphthalene	3.893	4.127	-6.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043608.D
Acq On : 29 Oct 2024 19:29
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 30 01:38:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.073	1.110	-3.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6