

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043586.D
 Acq On : 29 Oct 2024 10:57
 Operator : JC/MD
 Sample : VX1029WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024

Quant Time: Oct 30 01:26:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	202431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	352202	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	142936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	135908	42.105	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.220%		
35) Dibromofluoromethane	5.379	113	111394	46.642	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.280%		
50) Toluene-d8	8.647	98	399169	48.283	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.560%		
62) 4-Bromofluorobenzene	11.079	95	144111	47.042	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	94.080%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	40027	18.489	ug/l	100
3) Chloromethane	1.294	50	44310	16.012	ug/l	92
4) Vinyl Chloride	1.374	62	46678	18.174	ug/l	99
5) Bromomethane	1.593	94	18192	17.375	ug/l	97
6) Chloroethane	1.660	64	17156m	18.960	ug/l	
7) Trichlorofluoromethane	1.861	101	64804	20.654	ug/l	98
8) Diethyl Ether	2.136	74	25896	17.863	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.312	101	45384	21.388	ug/l	99
10) Methyl Iodide	2.440	142	57476	18.748	ug/l	97
11) Tert butyl alcohol	3.007	59	56407m	93.626	ug/l	
12) 1,1-Dichloroethene	2.306	96	42035	19.385	ug/l	92
13) Acrolein	2.239	56	40906	86.821	ug/l	99
14) Allyl chloride	2.654	41	68347	18.223	ug/l	96
15) Acrylonitrile	3.068	53	115840	85.162	ug/l	99
16) Acetone	2.386	43	118332	92.319	ug/l	98
17) Carbon Disulfide	2.501	76	82388	18.170	ug/l	100
18) Methyl Acetate	2.703	43	60142	16.054	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	143107	17.354	ug/l	100
20) Methylene Chloride	2.782	84	47149	17.068	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	44262	19.321	ug/l	97
22) Diisopropyl ether	3.757	45	139725	18.024	ug/l	92
23) Vinyl Acetate	3.721	43	571697	89.264	ug/l	99
24) 1,1-Dichloroethane	3.605	63	79799	17.909	ug/l	98
25) 2-Butanone	4.556	43	160914	86.062	ug/l	99
26) 2,2-Dichloropropane	4.465	77	72610	19.908	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	53987	18.448	ug/l	99
28) Bromochloromethane	4.891	49	37071	17.489	ug/l	96
29) Tetrahydrofuran	5.007	42	100228	82.487	ug/l	98
30) Chloroform	5.086	83	84330	18.055	ug/l	99
31) Cyclohexane	5.458	56	69310	20.338	ug/l	95
32) 1,1,1-Trichloroethane	5.373	97	76151	19.420	ug/l	99
36) 1,1-Dichloropropene	5.684	75	56285	19.875	ug/l	98
37) Ethyl Acetate	4.714	43	60171	17.501	ug/l	98
38) Carbon Tetrachloride	5.666	117	62906	20.104	ug/l	97
39) Methylcyclohexane	7.373	83	75030	21.340	ug/l	99
40) Benzene	6.031	78	178716	18.917	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
 Data File : VX043586.D
 Acq On : 29 Oct 2024 10:57
 Operator : JC/MD
 Sample : VX1029WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024

Quant Time: Oct 30 01:26:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	33159	18.374	ug/l	97
42) 1,2-Dichloroethane	6.080	62	62263	18.332	ug/l	99
43) Isopropyl Acetate	6.336	43	100112	17.607	ug/l	100
44) Trichloroethene	7.123	130	46116	19.649	ug/l	95
45) 1,2-Dichloropropane	7.427	63	42277	18.174	ug/l	100
46) Dibromomethane	7.580	93	32080	17.910	ug/l	96
47) Bromodichloromethane	7.818	83	57532	18.310	ug/l	99
48) Methyl methacrylate	7.696	41	48326	17.927	ug/l	98
49) 1,4-Dioxane	7.671	88	23130	417.634	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	312558	90.192	ug/l	99
52) Toluene	8.714	92	113851	19.871	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	59568	17.650	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	70181	19.517	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	44829	18.797	ug/l	93
56) Ethyl methacrylate	9.116	69	69497	17.934	ug/l	98
57) 1,3-Dichloropropane	9.305	76	74400	18.757	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	172110	91.703	ug/l	99
59) 2-Hexanone	9.433	43	240151	91.242	ug/l	100
60) Dibromochloromethane	9.518	129	44184	18.709	ug/l	98
61) 1,2-Dibromoethane	9.610	107	45897	18.539	ug/l	96
64) Tetrachloroethene	9.269	164	39340	21.213	ug/l	98
65) Chlorobenzene	10.079	112	123286	19.435	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	40788	19.685	ug/l	99
67) Ethyl Benzene	10.195	91	209043	20.209	ug/l	100
68) m/p-Xylenes	10.299	106	160547	40.586	ug/l	99
69) o-Xylene	10.640	106	79997	20.128	ug/l	99
70) Styrene	10.652	104	133590	20.223	ug/l	99
71) Bromoform	10.799	173	26501	17.788	ug/l #	96
73) Isopropylbenzene	10.963	105	207962	20.539	ug/l	100
74) N-amyl acetate	10.841	43	89068	18.058	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	67422	18.518	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	54362m	17.788	ug/l	
77) Bromobenzene	11.195	156	50119	19.170	ug/l	99
78) n-propylbenzene	11.305	91	227750	20.630	ug/l	99
79) 2-Chlorotoluene	11.360	91	143898	19.657	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	171034	20.260	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	19932	17.942	ug/l	96
82) 4-Chlorotoluene	11.451	91	162418	19.767	ug/l	96
83) tert-Butylbenzene	11.713	119	176326	20.537	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	176438	20.508	ug/l	99
85) sec-Butylbenzene	11.890	105	214438	21.308	ug/l	100
86) p-Isopropyltoluene	12.006	119	180325	21.080	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	92275	19.691	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	93923	19.343	ug/l	99
89) n-Butylbenzene	12.329	91	148112	21.249	ug/l	99
90) Hexachloroethane	12.536	117	28082	20.210	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	92312	19.265	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	12277	16.540	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	55722	19.158	ug/l	98
94) Hexachlorobutadiene	13.725	225	22819	21.279	ug/l	98
95) Naphthalene	13.774	128	207496	18.643	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	57807	18.850	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102924\
Data File : VX043586.D
Acq On : 29 Oct 2024 10:57
Operator : JC/MD
Sample : VX1029WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024

Quant Time: Oct 30 01:26:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Oct 30 01:26:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion:168 Resp: 20243
Ion Ratio Lower Upper
168 100
99 63.6 52.6 78.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#2
Dichlorodifluoromethane
Concen: 18.489 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 85 Resp: 40027
Ion Ratio Lower Upper
85 100
87 31.5 15.7 47.0









#3
Chloromethane
Concen: 16.012 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 50 Resp: 44310
Ion Ratio Lower Upper
50 100
52 37.2 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#4
Vinyl Chloride
Concen: 18.174 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 62 Resp: 46678
Ion Ratio Lower Upper
62 100
64 31.5 25.0 37.4









#5
Bromomethane
Concen: 17.375 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 94 Resp: 18192
Ion Ratio Lower Upper
94 100
96 97.6 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#6
Chloroethane
Concen: 18.960 ug/l m
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 64 Resp: 17156
Ion Ratio Lower Upper
64 100
66 30.3 25.1 37.7









#7
Trichlorofluoromethane
Concen: 20.654 ug/l
RT: 1.861 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion:101 Resp: 64804
Ion Ratio Lower Upper
101 100
103 65.4 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#8
Diethyl Ether
Concen: 17.863 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 74 Resp: 25896
Ion Ratio Lower Upper
74 100
45 96.5 47.5 142.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.388 ug/l
RT: 2.312 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion:101 Resp: 45384
Ion Ratio Lower Upper
101 100
85 42.1 34.5 51.7
151 73.1 58.2 87.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#10
Methyl Iodide
Concen: 18.748 ug/l
RT: 2.440 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion:142 Resp: 57476
Ion Ratio Lower Upper
142 100
127 45.5 38.2 57.4
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 93.626 ug/l m
RT: 3.007 min Scan# 31
Delta R.T. -0.043 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 59 Resp: 5640
Ion Ratio Lower Upper
59 100
57 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#12
1,1-Dichloroethene
Concen: 19.385 ug/l
RT: 2.306 min Scan# 200
Delta R.T. 0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 96 Resp: 42035
Ion Ratio Lower Upper
96 100
61 149.2 130.2 195.4
98 64.9 51.0 76.6





#13
Acrolein
Concen: 86.821 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 56 Resp: 40900
Ion Ratio Lower Upper
56 100
55 70.3 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#14
Allyl chloride
Concen: 18.223 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 41 Resp: 68347
Ion Ratio Lower Upper
41 100
39 63.1 53.2 79.8
76 35.3 27.0 40.6





#15

Acrylonitrile

Concen: 85.162 ug/l

RT: 3.068 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01





Tgt Ion: 53 Resp: 115840

Ion	Ratio	Lower	Upper
53	100		
52	82.5	65.8	98.6
51	36.4	28.5	42.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#16

Acetone

Concen: 92.319 ug/l

RT: 2.386 min Scan# 213




Delta R.T. -0.012 min


Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion: 43 Resp: 118332

Ion	Ratio	Lower	Upper
43	100		
58	32.4	25.1	37.7









#17
Carbon Disulfide
Concen: 18.170 ug/l
RT: 2.501 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 76 Resp: 8238
Ion Ratio Lower Upper
76 100
78 8.7 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#18
Methyl Acetate
Concen: 16.054 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 43 Resp: 60142
Ion Ratio Lower Upper
43 100
74 24.2 19.1 28.7





#19
Methyl tert-butyl Ether
Concen: 17.354 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 73 Resp: 14310
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#20
Methylene Chloride
Concen: 17.068 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 84 Resp: 47149
Ion Ratio Lower Upper
84 100
49 115.5 97.0 145.6
51 37.4 29.8 44.8
86 64.6 53.1 79.7





#21
trans-1,2-Dichloroethene
Concen: 19.321 ug/l
RT: 3.081 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 96 Resp: 4426
Ion Ratio Lower Upper
96 100
61 130.8 107.1 160.7
98 62.4 47.7 71.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#22
Diisopropyl ether
Concen: 18.024 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 45 Resp: 139725
Ion Ratio Lower Upper
45 100
43 85.7 60.6 90.8
87 28.5 22.0 33.0
59 12.1 9.5 14.3









#23
Vinyl Acetate
Concen: 89.264 ug/l
RT: 3.721 min Scan# 41
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 43 Resp: 57169
Ion Ratio Lower Upper
43 100
86 11.5 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#24
1,1-Dichloroethane
Concen: 17.909 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 63 Resp: 79799
Ion Ratio Lower Upper
63 100
98 8.0 3.5 10.4
100 5.2 2.4 7.2









#25
2-Butanone
Concen: 86.062 ug/l
RT: 4.556 min Scan# 571
Delta R.T. -0.012 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 43 Resp: 160914
Ion Ratio Lower Upper
43 100
72 26.7 21.1 31.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#26
2,2-Dichloropropane
Concen: 19.908 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 77 Resp: 72610
Ion Ratio Lower Upper
77 100
97 23.7 11.6 34.8





#27

cis-1,2-Dichloroethene

Concen: 18.448 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion: 96	Resp: 5398
Ion Ratio	Lower Upper
96 100	
61 139.6	0.0 277.2
98 64.1	0.0 131.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#28

Bromochloromethane

Concen: 17.489 ug/l




RT: 4.891 min Scan# 624


Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57


Tgt Ion: 49	Resp: 37071
Ion Ratio	Lower Upper
49 100	
129 2.3	0.0 4.2
130 88.6	67.8 101.8





#29
Tetrahydrofuran
Concen: 82.487 ug/l
RT: 5.007 min Scan# 644
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 42 Resp: 100228
Ion Ratio Lower Upper
42 100
72 47.2 36.6 55.0
71 43.8 34.2 51.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#30
Chloroform
Concen: 18.055 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 83 Resp: 84330
Ion Ratio Lower Upper
83 100
85 64.3 52.2 78.4





#31
Cyclohexane
Concen: 20.338 ug/l
RT: 5.458 min Scan# 717
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion: 56 Resp: 69310
Ion Ratio Lower Upper
56 100
69 33.4 24.9 37.3
84 86.4 73.3 109.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#32
1,1,1-Trichloroethane
Concen: 19.420 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57



Tgt Ion: 97 Resp: 76151
Ion Ratio Lower Upper
97 100
99 63.8 51.4 77.0
61 40.6 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 42.105 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01

Tgt Ion: 65 Resp: 135908

Ion Ratio Lower Upper

65 100





67 51.3 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57




Tgt Ion:114 Resp: 352202


Ion Ratio Lower Upper

114 100

63 19.6 0.0 41.2

88 15.6 0.0 34.0





#35

Dibromofluoromethane

Concen: 46.642 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:	113	Resp:	111394
Ion	Ratio	Lower	Upper
113	100		
111	101.7	81.4	122.0
192	18.3	14.1	21.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#36

1,1-Dichloropropene

Concen: 19.875 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion:	75	Resp:	56285
Ion	Ratio	Lower	Upper
75	100		
110	37.7	17.9	53.7
77	31.3	25.0	37.4









#37
Ethyl Acetate
Concen: 17.501 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 43 Resp: 6017
Ion Ratio Lower Upper
43 100
61 14.5 11.0 16.6
70 12.0 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#38
Carbon Tetrachloride
Concen: 20.104 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion:117 Resp: 62906
Ion Ratio Lower Upper
117 100
119 99.3 77.0 115.6
121 32.7 24.7 37.1





#39

Methylcyclohexane

Concen: 21.340 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01





Tgt Ion: 83 Resp: 75030

Ion	Ratio	Lower	Upper
83	100		
55	74.6	59.9	89.9
98	49.4	41.3	61.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#40

Benzene

Concen: 18.917 ug/l

RT: 6.031 min Scan# 811




Delta R.T. 0.000 min


Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion: 78 Resp: 178716

Ion	Ratio	Lower	Upper
78	100		
77	25.2	18.4	27.6









#41
Methacrylonitrile
Concen: 18.374 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 41	Resp: 3315
Ion Ratio	Lower Upper
41 100	
39 54.2	43.5 65.3
67 77.0	58.3 87.5
52 30.3	24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#42
1,2-Dichloroethane
Concen: 18.332 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 62	Resp: 62263
Ion Ratio	Lower Upper
62 100	
98 9.9	0.0 18.8





#43
Isopropyl Acetate
Concen: 17.607 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 43 Resp: 100112
Ion Ratio Lower Upper
43 100
61 20.5 16.2 24.4
87 14.0 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#44
Trichloroethene
Concen: 19.649 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion:130 Resp: 46116
Ion Ratio Lower Upper
130 100
95 96.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 18.174 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01

Tgt Ion: 63 Resp: 4227

Ion Ratio Lower Upper

63 100





65 30.8 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#46

Dibromomethane

Concen: 17.910 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57




Tgt Ion: 93 Resp: 32080


Ion Ratio Lower Upper

93 100

95 85.9 66.4 99.6

174 98.8 75.0 112.4









#47
Bromodichloromethane
Concen: 18.310 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01

Tgt Ion: 83 Resp: 5753
Ion Ratio Lower Upper
83 100
85 65.7 52.2 78.4
127 10.0 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#48
Methyl methacrylate
Concen: 17.927 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 41 Resp: 48326
Ion Ratio Lower Upper
41 100
69 85.0 67.7 101.5
39 54.6 46.3 69.5





#49

1,4-Dioxane

Concen: 417.634 ug/l

RT: 7.671 min Scan# 1080

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion: 88	Resp: 23130
Ion Ratio	Lower Upper
88 100	
43 31.2	25.3 37.9
58 64.9	53.9 80.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#50

Toluene-d8

Concen: 48.283 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57


Tgt Ion: 98	Resp: 399169
Ion Ratio	Lower Upper
98 100	
100 67.1	53.4 80.2





#51
4-Methyl-2-Pentanone
Concen: 90.192 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 43 Resp: 31255
Ion Ratio Lower Upper
43 100
58 39.0 31.7 47.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#52
Toluene
Concen: 19.871 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Tgt Ion: 92 Resp: 113851
Ion Ratio Lower Upper
92 100
91 169.0 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 17.650 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01

Tgt Ion: 75 Resp: 5956

Ion Ratio Lower Upper

75 100





77 31.1 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#54

cis-1,3-Dichloropropene

Concen: 19.517 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57




Tgt Ion: 75 Resp: 70181


Ion Ratio Lower Upper

75 100

77 31.4 24.8 37.2


39 43.5 37.7 56.5





#55
1,1,2-Trichloroethane
Concen: 18.797 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01






Tgt Ion: 97 Resp: 44829


Ion	Ratio	Lower	Upper
97	100		
83	75.7	67.3	100.9
85	50.3	43.6	65.4
99	58.7	50.6	75.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024






#56
Ethyl methacrylate
Concen: 17.934 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57



Tgt Ion: 69 Resp: 69497

Ion	Ratio	Lower	Upper
69	100		
41	65.7	53.8	80.8
39	35.5	29.9	44.9





#57

1,3-Dichloropropane

Concen: 18.757 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01

Tgt Ion: 76 Resp: 74400

Ion Ratio Lower Upper

76 100





78 33.3 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#58

2-Chloroethyl Vinyl ether

Concen: 91.703 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX043586.D




Acq: 29 Oct 2024 10:57


Tgt Ion: 63 Resp: 172110

Ion Ratio Lower Upper

63 100


106 30.2 23.6 35.4





#59
2-Hexanone
Concen: 91.242 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 43 Resp: 24015
Ion Ratio Lower Upper
43 100
58 54.8 27.5 82.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#60
Dibromochloromethane
Concen: 18.709 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion:129 Resp: 44184
Ion Ratio Lower Upper
129 100
127 74.8 38.4 115.1





#61
1,2-Dibromoethane
Concen: 18.539 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion:107 Resp: 4589
Ion Ratio Lower Upper
107 100
109 92.8 77.1 115.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#62
4-Bromofluorobenzene
Concen: 47.042 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Tgt Ion: 95 Resp: 144111
Ion Ratio Lower Upper
95 100
174 75.5 0.0 146.6
176 72.4 0.0 139.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion:117 Resp: 29485
Ion Ratio Lower Upper
117 100
82 53.1 42.1 63.1
119 33.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#64
Tetrachloroethene
Concen: 21.213 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion:164 Resp: 39340
Ion Ratio Lower Upper
164 100
166 130.7 103.8 155.6
129 101.0 83.8 125.6
131 97.8 80.6 121.0





#65
Chlorobenzene
Concen: 19.435 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion:112 Resp: 123280
Ion Ratio Lower Upper
112 100
114 31.6 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 19.685 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Tgt Ion:131 Resp: 40788
Ion Ratio Lower Upper
131 100
133 96.2 48.0 144.0
119 66.7 32.3 96.9





#67
Ethyl Benzene
Concen: 20.209 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion: 91 Resp: 20904
Ion Ratio Lower Upper
91 100
106 32.1 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#68
m/p-Xylenes
Concen: 40.586 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57



Tgt Ion:106 Resp: 160547
Ion Ratio Lower Upper
106 100
91 204.3 164.7 247.1





#69

o-Xylene

Concen: 20.128 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:106 Resp: 7999
 Ion Ratio Lower Upper
 106 100
 91 213.4 107.9 323.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#70

Styrene

Concen: 20.223 ug/l




RT: 10.652 min Scan# 1569


Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion:104 Resp: 133590
 Ion Ratio Lower Upper
 104 100
 78 49.8 40.0 60.0
 103 53.3 43.4 65.0





#71

Bromoform

Concen: 17.788 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:173 Resp: 2650
 Ion Ratio Lower Upper
 173 100
 175 49.1 23.3 69.9
 254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57


Tgt Ion:152 Resp: 142936
 Ion Ratio Lower Upper
 152 100
 115 69.7 42.9 128.7
 150 160.8 0.0 343.6





#73
Isopropylbenzene
Concen: 20.539 ug/l
RT: 10.963 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion:105 Resp: 207962
Ion Ratio Lower Upper
105 100
120 26.5 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#74
N-ethyl acetate
Concen: 18.058 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Tgt Ion: 43 Resp: 89068
Ion Ratio Lower Upper
43 100
70 44.0 34.6 52.0
55 24.4 19.6 29.4
61 24.6 20.2 30.2





#75
1,1,2,2-Tetrachloroethane
Concen: 18.518 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion: 83 Resp: 6742
Ion Ratio Lower Upper
83 100
131 10.1 5.3 15.9
85 63.4 32.4 97.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#76
1,2,3-Trichloropropane
Concen: 17.788 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57



Tgt Ion: 75 Resp: 54362
Ion Ratio Lower Upper
75 100
77 45.1 22.4 67.2





#77

Bromobenzene

Concen: 19.170 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:	156	Resp:	50119
Ion Ratio	Lower	Upper	
156	100		
77	157.9	79.7	239.1
158	96.1	48.0	144.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#78

n-propylbenzene

Concen: 20.630 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion:	91	Resp:	227750
Ion Ratio	Lower	Upper	
91	100		
120	23.5	11.6	34.7





#79

2-Chlorotoluene

Concen: 19.657 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01

Tgt Ion: 91 Resp: 143898

Ion Ratio Lower Upper

91 100





126 34.0 16.7 50.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#80

1,3,5-Trimethylbenzene

Concen: 20.260 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX043586.D




Acq: 29 Oct 2024 10:57


Tgt Ion:105 Resp: 171034

Ion Ratio Lower Upper

105 100


120 49.4 23.8 71.5





#81
trans-1,4-Dichloro-2-butene
Concen: 17.942 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 75 Resp: 19932
Ion Ratio Lower Upper
75 100
53 101.5 84.6 126.8
89 45.9 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#82
4-Chlorotoluene
Concen: 19.767 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion: 91 Resp: 162418
Ion Ratio Lower Upper
91 100
126 30.5 14.2 42.8





#83
tert-Butylbenzene
Concen: 20.537 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion:119 Resp: 176320
Ion Ratio Lower Upper
119 100
91 59.0 29.5 88.5
134 23.8 11.8 35.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#84
1,2,4-Trimethylbenzene
Concen: 20.508 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Tgt Ion:105 Resp: 176438
Ion Ratio Lower Upper
105 100
120 43.5 22.1 66.1





#85
 sec-Butylbenzene
 Concen: 21.308 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX043586.D
 Acq: 29 Oct 2024 10:57




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029WBS01




Tgt Ion:105 Resp: 214438
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.8 29.5

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024




#86
 p-Isopropyltoluene
 Concen: 21.080 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX043586.D
 Acq: 29 Oct 2024 10:57



Tgt Ion:119 Resp: 180325
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.1 39.1
 91 24.2 12.4 37.0





#87

1,3-Dichlorobenzene

Concen: 19.691 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX043586.D


Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :

VX1029WBS01





Tgt Ion:146 Resp: 92275

Ion	Ratio	Lower	Upper
146	100		
111	41.3	21.1	63.4
148	63.1	31.6	94.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#88

1,4-Dichlorobenzene


Concen: 19.343 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min



Lab File: VX043586.D


Acq: 29 Oct 2024 10:57



Tgt Ion:146 Resp: 93923


Ion	Ratio	Lower	Upper
146	100		
111	40.1	20.4	61.3
148	64.4	31.9	95.9





#89
n-Butylbenzene
Concen: 21.249 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01



Tgt Ion: 91 Resp: 14811
Ion Ratio Lower Upper
91 100
92 55.1 27.1 81.2
134 26.1 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#90
Hexachloroethane
Concen: 20.210 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57


Tgt Ion:117 Resp: 28082
Ion Ratio Lower Upper
117 100
201 65.0 32.9 98.7





#91
1,2-Dichlorobenzene
Concen: 19.265 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57




Instrument :
MSVOA_X
ClientSampleId :
VX1029WBS01




Tgt Ion:146 Resp: 92312
Ion Ratio Lower Upper
146 100
111 43.5 21.6 64.6
148 62.9 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 16.540 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX043586.D
Acq: 29 Oct 2024 10:57



Tgt Ion: 75 Resp: 12277
Ion Ratio Lower Upper
75 100
155 85.4 40.1 120.3
157 107.8 50.8 152.4





#93

1,2,4-Trichlorobenzene

Concen: 19.158 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:	180	Resp:	5572
Ion Ratio	Lower	Upper	
180	100		
182	95.6	46.8	140.4
145	33.8	16.1	48.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#94

Hexachlorobutadiene

Concen: 21.279 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion:	225	Resp:	22819
Ion Ratio	Lower	Upper	
225	100		
223	63.2	30.8	92.4
227	62.5	31.6	94.8





#95

Naphthalene

Concen: 18.643 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Instrument :

MSVOA_X

ClientSampleId :





VX1029WBS01

Tgt Ion:128 Resp: 207490
 Ion Ratio Lower Upper
 128 100
 127 12.7 10.2 15.4
 129 11.1 8.7 13.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#96

1,2,3-Trichlorobenzene

Concen: 18.850 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.006 min

Lab File: VX043586.D

Acq: 29 Oct 2024 10:57

Tgt Ion:180 Resp: 57807
 Ion Ratio Lower Upper
 180 100
 182 96.0 47.1 141.3
 145 34.6 17.4 52.3

