

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043612.D
 Acq On : 30 Oct 2024 10:22
 Operator : JC/MD
 Sample : VX1030WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1030WBL01

Quant Time: Oct 31 09:41:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	124305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	229429	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	87309	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86399	43.590	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.180%
35) Dibromofluoromethane	5.385	113	70244	45.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.300%
50) Toluene-d8	8.647	98	266605	49.505	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	93954	47.081	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043612.D
Acq On : 30 Oct 2024 10:22
Operator : JC/MD
Sample : VX1030WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1030WBL01

Quant Time: Oct 31 09:41:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

