

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043616.D
 Acq On : 30 Oct 2024 12:06
 Operator : JC/MD
 Sample : P4600-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190-GW-478-480

Quant Time: Oct 30 13:06:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	259096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95599	41.880	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.760%
35) Dibromofluoromethane	5.385	113	79955	45.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	304017	49.988	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.980%
62) 4-Bromofluorobenzene	11.079	95	106465	47.242	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.480%
Target Compounds						
16) Acetone	2.386	43	3065	3.381	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043616.D
Acq On : 30 Oct 2024 12:06
Operator : JC/MD
Sample : P4600-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190-GW-478-480

Quant Time: Oct 30 13:06:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

