

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043618.D
 Acq On : 30 Oct 2024 12:52
 Operator : JC/MD
 Sample : P4562-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Oct 31 09:42:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	256051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92697	41.456	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.920%		
35) Dibromofluoromethane	5.385	113	77720	44.763	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.520%		
50) Toluene-d8	8.647	98	301484	50.161	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.320%		
62) 4-Bromofluorobenzene	11.079	95	109421	49.131	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.260%		
Target Compounds						
16) Acetone	2.392	43	1587	1.787	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043618.D
Acq On : 30 Oct 2024 12:52
Operator : JC/MD
Sample : P4562-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Oct 31 09:42:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

