

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043620.D
 Acq On : 30 Oct 2024 13:38
 Operator : JC/MD
 Sample : P4615-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-10292024

Quant Time: Oct 31 09:42:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	141860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92963	41.098	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.200%		
35) Dibromofluoromethane	5.385	113	76925	44.017	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.040%		
50) Toluene-d8	8.647	98	300447	49.664	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.320%		
62) 4-Bromofluorobenzene	11.079	95	106618	47.561	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.120%		
Target Compounds						
16) Acetone	2.392	43	1556	1.732	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043620.D
Acq On : 30 Oct 2024 13:38
Operator : JC/MD
Sample : P4615-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-10292024

Quant Time: Oct 31 09:42:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

