

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043621.D
 Acq On : 30 Oct 2024 14:01
 Operator : JC/MD
 Sample : P4547-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F-21

Quant Time: Oct 31 09:42:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	250365	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91456	42.214	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.420%
35) Dibromofluoromethane	5.379	113	75902	44.708	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.420%
50) Toluene-d8	8.647	98	292278	49.734	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.460%
62) 4-Bromofluorobenzene	11.079	95	107444	49.339	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.680%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	36607	42.550	ug/l	94
20) Methylene Chloride	2.782	84	2771	1.494	ug/l	86
25) 2-Butanone	4.593	43	3915	3.120	ug/l #	90
43) Isopropyl Acetate	6.342	43	110523	27.345	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043621.D
Acq On : 30 Oct 2024 14:01
Operator : JC/MD
Sample : P4547-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F-21

Quant Time: Oct 31 09:42:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

