

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
 Data File : VX043631.D
 Acq On : 30 Oct 2024 17:53
 Operator : JC/MD
 Sample : PB164501ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB164501ZHE#09

Quant Time: Oct 31 23:03:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	151985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	279240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98645	40.705	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	81.400%		
35) Dibromofluoromethane	5.385	113	83824	44.269	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.540%		
50) Toluene-d8	8.647	98	319095	48.682	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.360%		
62) 4-Bromofluorobenzene	11.079	95	106964	44.039	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	88.080%		
Target Compounds						
16) Acetone	2.392	43	29242	30.386	ug/l	98
18) Methyl Acetate	2.709	43	3658	1.301	ug/l	98
20) Methylene Chloride	2.782	84	2763	1.332	ug/l	92
43) Isopropyl Acetate	6.342	43	123624	27.424	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103024\
Data File : VX043631.D
Acq On : 30 Oct 2024 17:53
Operator : JC/MD
Sample : PB164501ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB164501ZHE#09

Quant Time: Oct 31 23:03:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

