

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103119\
 Data File : VX013298.D
 Acq On : 31 Oct 2019 18:39
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX103119

Quant Time: Nov 02 06:11:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.576	0.599	-4.0	102	0.00
3 P	Chloromethane	0.625	0.635	-1.6	101	0.00
4 C	Vinyl Chloride	0.711	0.742	-4.4#	101	0.00
5 T	Bromomethane	0.496	0.517	-4.2	107	0.00
6 T	Chloroethane	0.454	0.405	10.8	101	0.00
7 T	Trichlorofluoromethane	0.840	0.825	1.8	102	0.00
8 T	Diethyl Ether	0.372	0.364	2.2	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.471	-0.6	105	0.00
10 T	Methyl Iodide	0.580	0.577	0.5	98	0.00
11 T	Tert butyl alcohol	0.156	0.144	7.7	102	0.00
12 CM	1,1-Dichloroethene	0.459	0.470	-2.4#	103	0.00
13 T	Acrolein	0.085	0.086	-1.2	101	0.00
14 T	Allyl chloride	0.833	0.849	-1.9	103	0.00
15 T	Acrylonitrile	0.292	0.303	-3.8	102	0.00
16 T	Acetone	0.295	0.284	3.7	96	0.00
17 T	Carbon Disulfide	1.228	1.317	-7.2	108	0.00
18 T	Methyl Acetate	0.816	0.840	-2.9	102	0.00
19 T	Methyl tert-butyl Ether	1.530	1.576	-3.0	102	0.00
20 T	Methylene Chloride	0.546	0.521	4.6	102	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.501	-0.4	102	0.00
22 T	Diisopropyl ether	1.609	1.650	-2.5	101	0.00
23 T	Vinyl Acetate	1.377	1.428	-3.7	102	0.00
24 P	1,1-Dichloroethane	0.900	0.907	-0.8	103	0.00
25 T	2-Butanone	0.433	0.438	-1.2	98	0.00
26 T	2,2-Dichloropropane	0.716	0.715	0.1	101	0.00
27 T	cis-1,2-Dichloroethene	0.559	0.567	-1.4	101	0.00
28 T	Bromochloromethane	0.256	0.261	-2.0	103	0.00
29 T	Tetrahydrofuran	0.267	0.277	-3.7	99	0.00
30 C	Chloroform	0.889	0.895	-0.7#	102	0.00
31 T	Cyclohexane	0.810	0.817	-0.9	104	0.00
32 T	1,1,1-Trichloroethane	0.765	0.787	-2.9	104	0.00
33 S	1,2-Dichloroethane-d4	0.592	0.582	1.7	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.314	0.313	0.3	103	0.00
36 T	1,1-Dichloropropene	0.429	0.445	-3.7	102	0.00
37 T	Ethyl Acetate	0.530	0.547	-3.2	102	0.00
38 T	Carbon Tetrachloride	0.419	0.444	-6.0	104	0.00
39 T	Methylcyclohexane	0.534	0.546	-2.2	105	0.00
40 TM	Benzene	1.336	1.342	-0.4	102	0.00
41 T	Methacrylonitrile	0.297	0.299	-0.7	101	0.00
42 TM	1,2-Dichloroethane	0.482	0.495	-2.7	101	0.00
43 T	Isopropyl Acetate	0.825	0.860	-4.2	101	0.00
44 TM	Trichloroethene	0.368	0.368	0.0	103	0.00
45 C	1,2-Dichloropropane	0.340	0.345	-1.5#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103119\
 Data File : VX013298.D
 Acq On : 31 Oct 2019 18:39
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX103119

Quant Time: Nov 02 06:11:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.232	0.232	0.0	101	0.00
47 T	Bromodichloromethane	0.438	0.456	-4.1	102	0.00
48 T	Methyl methacrylate	0.412	0.434	-5.3	100	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	97	0.00
50 S	Toluene-d8	1.182	1.192	-0.8	102	0.00
51 T	4-Methyl-2-Pentanone	0.541	0.556	-2.8	99	0.00
52 CM	Toluene	0.828	0.846	-2.2#	102	0.00
53 T	t-1,3-Dichloropropene	0.479	0.519	-8.4	103	0.00
54 T	cis-1,3-Dichloropropene	0.536	0.566	-5.6	103	0.00
55 T	1,1,2-Trichloroethane	0.343	0.346	-0.9	101	0.00
56 T	Ethyl methacrylate	0.530	0.570	-7.5	103	0.00
57 T	1,3-Dichloropropane	0.583	0.582	0.2	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.235	0.251	-6.8	101	0.00
59 T	2-Hexanone	0.418	0.435	-4.1	99	0.00
60 T	Dibromochloromethane	0.351	0.377	-7.4	103	0.00
61 T	1,2-Dibromoethane	0.364	0.372	-2.2	101	0.00
62 S	4-Bromofluorobenzene	0.427	0.422	1.2	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.358	0.357	0.3	102	0.00
65 PM	Chlorobenzene	1.012	1.010	0.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.385	-3.5	102	0.00
67 C	Ethyl Benzene	1.747	1.798	-2.9#	101	0.00
68 T	m/p-Xylenes	0.661	0.681	-3.0	100	0.00
69 T	o-Xylene	0.647	0.670	-3.6	102	0.00
70 T	Styrene	1.081	1.150	-6.4	102	0.00
71 P	Bromoform	0.300	0.329	-9.7	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.315	3.429	-3.4	102	0.00
74 T	N-amyl acetate	1.537	1.674	-8.9	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.209	1.213	-0.3	100	0.00
76 T	1,2,3-Trichloropropane	1.016	0.971	4.4	100	0.00
77 T	Bromobenzene	0.884	0.903	-2.1	102	0.00
78 T	n-propylbenzene	3.693	3.885	-5.2	101	0.00
79 T	2-Chlorotoluene	2.225	2.282	-2.6	101	0.00
80 T	1,3,5-Trimethylbenzene	2.761	2.889	-4.6	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.395	0.413	-4.6	100	0.00
82 T	4-Chlorotoluene	2.604	2.653	-1.9	101	0.00
83 T	tert-Butylbenzene	2.780	2.882	-3.7	99	0.00
84 T	1,2,4-Trimethylbenzene	2.770	2.883	-4.1	100	0.00
85 T	sec-Butylbenzene	3.195	3.349	-4.8	102	0.00
86 T	p-Isopropyltoluene	2.942	3.090	-5.0	101	0.00
87 T	1,3-Dichlorobenzene	1.575	1.590	-1.0	100	0.00
88 T	1,4-Dichlorobenzene	1.605	1.613	-0.5	102	0.00
89 T	n-Butylbenzene	2.567	2.716	-5.8	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103119\
Data File : VX013298.D
Acq On : 31 Oct 2019 18:39
Operator : JC/SP
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX103119

Quant Time: Nov 02 06:11:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.524	0.568	-8.4	105	0.00
91 T	1,2-Dichlorobenzene	1.575	1.601	-1.7	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.290	1.7	100	0.00
93 T	1,2,4-Trichlorobenzene	1.098	1.167	-6.3	105	0.00
94 T	Hexachlorobutadiene	0.537	0.546	-1.7	103	0.00
95 T	Naphthalene	3.446	3.772	-9.5	101	0.00
96 T	1,2,3-Trichlorobenzene	1.102	1.151	-4.4	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6