

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043638.D
 Acq On : 31 Oct 2024 09:50
 Operator : JC/MD
 Sample : VX1031MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1031MBL01

Quant Time: Nov 01 00:47:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	264587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227412	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	99285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95093	40.257	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.520%
35) Dibromofluoromethane	5.379	113	80659	44.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.920%
50) Toluene-d8	8.647	98	306686	49.380	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.760%
62) 4-Bromofluorobenzene	11.079	95	105496	45.840	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043638.D
Acq On : 31 Oct 2024 09:50
Operator : JC/MD
Sample : VX1031MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1031MBL01

Quant Time: Nov 01 00:47:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

