

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043639.D
 Acq On : 31 Oct 2024 10:13
 Operator : JC/MD
 Sample : VX1031WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1031WBL01

Quant Time: Nov 01 00:47:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	263897	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92853	39.986	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.980%
35) Dibromofluoromethane	5.379	113	81033	45.283	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.560%
50) Toluene-d8	8.647	98	300250	48.470	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.079	95	101416	44.183	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043639.D
Acq On : 31 Oct 2024 10:13
Operator : JC/MD
Sample : VX1031WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1031WBL01

Quant Time: Nov 01 00:47:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

