

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043642.D
 Acq On : 31 Oct 2024 11:37
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Nov 01 00:50:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	143345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	264743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108803	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93492	40.904	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.800%
35) Dibromofluoromethane	5.379	113	79550	44.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.620%
50) Toluene-d8	8.646	98	306348	49.297	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	108980	47.326	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043642.D
Acq On : 31 Oct 2024 11:37
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Nov 01 00:50:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

