

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043645.D
 Acq On : 31 Oct 2024 12:54
 Operator : JC/MD
 Sample : P4616-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F9-MOVED

Quant Time: Nov 01 00:51:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	203439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	179720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79334	45.062	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.120%
35) Dibromofluoromethane	5.385	113	64075	46.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.647	98	238398	49.922	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.840%
62) 4-Bromofluorobenzene	11.079	95	82392	46.562	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.120%
Target Compounds						
16) Acetone	2.386	43	33169	47.444	ug/l	99
20) Methylene Chloride	2.782	84	2264	1.503	ug/l #	84
43) Isopropyl Acetate	6.342	43	94139	28.664	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043645.D
Acq On : 31 Oct 2024 12:54
Operator : JC/MD
Sample : P4616-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F9-MOVED

Quant Time: Nov 01 00:51:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

