

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043646.D
 Acq On : 31 Oct 2024 13:17
 Operator : JC/MD
 Sample : P4617-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-PILE

Quant Time: Nov 01 00:52:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111169	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	203279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76177	42.974	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.940%
35) Dibromofluoromethane	5.385	113	53119	38.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	77.080%
50) Toluene-d8	8.647	98	240873	50.480	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	84536	47.811	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	39120	55.575	ug/l	99
18) Methyl Acetate	2.709	43	8218	3.995	ug/l	98
20) Methylene Chloride	2.782	84	1621	1.069	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043646.D
Acq On : 31 Oct 2024 13:17
Operator : JC/MD
Sample : P4617-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-PILE

Quant Time: Nov 01 00:52:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

