

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
 Data File : VX043650.D
 Acq On : 31 Oct 2024 14:50
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Nov 01 00:54:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	199063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178625	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73111	41.145	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.280%
35) Dibromofluoromethane	5.385	113	60047	44.485	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.960%
50) Toluene-d8	8.647	98	227990	48.793	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	82188	47.468	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.940%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103124\
Data File : VX043650.D
Acq On : 31 Oct 2024 14:50
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Nov 01 00:54:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

