

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110124\
 Data File : VX043655.D
 Acq On : 01 Nov 2024 10:52
 Operator : JC/MD
 Sample : VX1101MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1101MBL01

Quant Time: Nov 04 06:08:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:41:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	140092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	253243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	224850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	103468	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	87778	39.295	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.600%
35) Dibromofluoromethane	5.373	113	76336	44.453	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.900%
50) Toluene-d8	8.647	98	296427	49.866	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.079	95	106060	48.150	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110124\
Data File : VX043655.D
Acq On : 01 Nov 2024 10:52
Operator : JC/MD
Sample : VX1101MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1101MBL01

Quant Time: Nov 04 06:08:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:41:34 2024
Response via : Initial Calibration

