

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110124\  
 Data File : VX043656.D  
 Acq On : 01 Nov 2024 11:15  
 Operator : JC/MD  
 Sample : VX1101WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1101WBL01

Quant Time: Nov 04 06:09:19 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102824W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Oct 28 13:41:34 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 146410   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 269747   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.049 | 117   | 232441   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 102414   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 95733    | 41.007   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 82.020%  |
| 35) Dibromofluoromethane    | 5.379  | 113   | 82312    | 45.000   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 90.000%  |
| 50) Toluene-d8              | 8.647  | 98    | 307422   | 48.552   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 97.100%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 105720   | 45.059   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =     | 90.120%  |

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX110124\  
Data File : VX043656.D  
Acq On : 01 Nov 2024 11:15  
Operator : JC/MD  
Sample : VX1101WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1101WBL01

Quant Time: Nov 04 06:09:19 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102824W.M  
Quant Title : SW846 8260  
QLast Update : Mon Oct 28 13:41:34 2024  
Response via : Initial Calibration

