

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032571.D
 Acq On : 04 Nov 2022 13:09
 Operator : JC/MD
 Sample : N5204-01ME
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20071ME

Quant Time: Nov 04 13:33:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	116632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78821	48.638	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.280%		
35) Dibromofluoromethane	5.391	113	59984	43.016	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	86.040%		
50) Toluene-d8	8.646	98	229693	47.371	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.740%		
62) 4-Bromofluorobenzene	11.085	95	77462	40.897	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	81.800%#		
Target Compounds						
16) Acetone	2.410	43	10183	17.289	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032571.D
Acq On : 04 Nov 2022 13:09
Operator : JC/MD
Sample : N5204-01ME
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20071ME

Quant Time: Nov 04 13:33:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

