

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032572.D
 Acq On : 04 Nov 2022 13:32
 Operator : JC/MD
 Sample : N5395-07ME
 Misc : 5.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-3-103122

Quant Time: Nov 05 00:36:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194475	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80039	48.923	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.840%
35) Dibromofluoromethane	5.391	113	60521	43.628	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.260%
50) Toluene-d8	8.647	98	230579	47.802	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.600%
62) 4-Bromofluorobenzene	11.085	95	73595	39.059	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.120%#
Target Compounds						
52) Toluene	8.720	92	8422	2.361	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032572.D
Acq On : 04 Nov 2022 13:32
Operator : JC/MD
Sample : N5395-07ME
Misc : 5.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-3-103122

Quant Time: Nov 05 00:36:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

