

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032573.D
 Acq On : 04 Nov 2022 13:55
 Operator : JC/MD
 Sample : N5395-13ME
 Misc : 5.83g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-13-103122

Quant Time: Nov 05 00:36:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196394	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80349	49.109	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.220%
35) Dibromofluoromethane	5.385	113	59274	42.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.620%
50) Toluene-d8	8.653	98	233626	47.960	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.920%
62) 4-Bromofluorobenzene	11.079	95	73065	38.398	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	76.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032573.D
Acq On : 04 Nov 2022 13:55
Operator : JC/MD
Sample : N5395-13ME
Misc : 5.83g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-13-103122

Quant Time: Nov 05 00:36:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

