

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032585.D
 Acq On : 04 Nov 2022 18:34
 Operator : JC/MD
 Sample : N5486-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Nov 05 00:39:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77599	49.115	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.240%
35) Dibromofluoromethane	5.391	113	60027	45.021	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.040%
50) Toluene-d8	8.653	98	222894	48.077	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.160%
62) 4-Bromofluorobenzene	11.079	95	84244	46.518	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.040%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1411	2.457	ug/l #	86
19) Methyl tert-butyl Ether	3.111	73	155010	38.676	ug/l #	87
22) Diisopropyl ether	3.757	45	106284	27.032	ug/l #	86
24) 1,1-Dichloroethane	3.611	63	2497	1.104	ug/l #	87
27) cis-1,2-Dichloroethene	4.489	96	2018	1.407	ug/l	96
31) Cyclohexane	5.483	56	6625	3.458	ug/l #	93
39) Methylcyclohexane	7.379	83	3845	1.750	ug/l #	78
40) Benzene	6.044	78	9626	1.832	ug/l	97
67) Ethyl Benzene	10.195	91	4560	0.683	ug/l #	86
68) m/p-Xylenes	10.305	106	1402	0.540	ug/l	96
73) Isopropylbenzene	10.963	105	3284	0.540	ug/l	97
78) n-propylbenzene	11.305	91	5595	0.799	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032585.D
Acq On : 04 Nov 2022 18:34
Operator : JC/MD
Sample : N5486-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Nov 05 00:39:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

