

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032598.D
 Acq On : 05 Nov 2022 02:27
 Operator : JC/MD
 Sample : N5483-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Nov 07 05:07:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85886	49.774	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.540%
35) Dibromofluoromethane	5.391	113	66676	45.474	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.940%
50) Toluene-d8	8.653	98	247700	48.583	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	98371	49.393	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.780%
Target Compounds						
16) Acetone	2.386	43	2182	3.479	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032598.D
Acq On : 05 Nov 2022 02:27
Operator : JC/MD
Sample : N5483-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Nov 07 05:07:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

