

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032605.D
 Acq On : 05 Nov 2022 05:10
 Operator : JC/MD
 Sample : N5483-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9R

Quant Time: Nov 07 05:09:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188647	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77443	49.111	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.220%
35) Dibromofluoromethane	5.385	113	60685	45.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.200%
50) Toluene-d8	8.653	98	225280	48.146	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.300%
62) 4-Bromofluorobenzene	11.079	95	84322	46.134	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.260%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1256	2.191	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032605.D
Acq On : 05 Nov 2022 05:10
Operator : JC/MD
Sample : N5483-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9R

Quant Time: Nov 07 05:09:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

