

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032607.D
 Acq On : 05 Nov 2022 05:56
 Operator : JC/MD
 Sample : N5484-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Nov 07 05:09:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78555	49.541	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.080%
35) Dibromofluoromethane	5.385	113	61339	45.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.740%
50) Toluene-d8	8.653	98	226736	48.229	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.460%
62) 4-Bromofluorobenzene	11.079	95	86757	47.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032607.D
Acq On : 05 Nov 2022 05:56
Operator : JC/MD
Sample : N5484-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Nov 07 05:09:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

