

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110422\
 Data File : VX032589.D
 Acq On : 04 Nov 2022 20:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.577	0.633	-9.7	96	0.00
3 P	Chloromethane	0.488	0.510	-4.5	95	0.00
4 C	Vinyl Chloride	0.537	0.580	-8.0#	96	0.00
5 T	Bromomethane	0.429	0.352	17.9	74	0.00
6 T	Chloroethane	0.360	0.378	-5.0	103	0.00
7 T	Trichlorofluoromethane	1.017	1.046	-2.9	95	0.00
8 T	Diethyl Ether	0.330	0.349	-5.8	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.561	-4.5	100	0.00
10 T	Methyl Iodide	0.558	0.500	10.4	78	0.00
11 T	Tert butyl alcohol	0.101	0.097	4.0	91	0.01
12 CM	1,1-Dichloroethene	0.501	0.518	-3.4#	97	0.00
13 T	Acrolein	0.101	0.117	-15.8	113	0.00
14 T	Allyl chloride	0.799	0.837	-4.8	97	0.00
15 T	Acrylonitrile	0.274	0.278	-1.5	99	0.00
16 T	Acetone	0.253	0.251	0.8	101	0.00
17 T	Carbon Disulfide	1.194	1.028	13.9	78	0.00
18 T	Methyl Acetate	0.821	0.905	-10.2	104	0.00
19 T	Methyl tert-butyl Ether	1.762	1.889	-7.2	103	0.00
20 T	Methylene Chloride	0.579	0.584	-0.9	101	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.548	-0.4	96	0.00
22 T	Diisopropyl ether	1.729	1.840	-6.4	103	0.00
23 T	Vinyl Acetate	1.397	1.426	-2.1	96	0.00
24 P	1,1-Dichloroethane	0.994	1.045	-5.1	100	0.00
25 T	2-Butanone	0.391	0.381	2.6	95	0.00
26 T	2,2-Dichloropropane	0.791	0.747	5.6	88	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.660	-4.6	98	0.00
28 T	Bromochloromethane	0.354	0.338	4.5	95	0.00
29 T	Tetrahydrofuran	0.249	0.243	2.4	93	0.00
30 C	Chloroform	1.091	1.151	-5.5#	101	0.00
31 T	Cyclohexane	0.842	0.848	-0.7	93	0.00
32 T	1,1,1-Trichloroethane	0.979	1.027	-4.9	98	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.701	-0.9	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.357	0.362	-1.4	98	0.00
36 T	1,1-Dichloropropene	0.497	0.510	-2.6	97	0.00
37 T	Ethyl Acetate	0.517	0.487	5.8	95	0.00
38 T	Carbon Tetrachloride	0.551	0.556	-0.9	94	0.00
39 T	Methylcyclohexane	0.588	0.550	6.5	85	0.00
40 TM	Benzene	1.405	1.437	-2.3	97	0.00
41 T	Methacrylonitrile	0.263	0.285	-8.4	103	0.00
42 TM	1,2-Dichloroethane	0.575	0.602	-4.7	99	0.00
43 T	Isopropyl Acetate	0.794	0.803	-1.1	94	0.00
44 TM	Trichloroethene	0.396	0.398	-0.5	97	0.00
45 C	1,2-Dichloropropane	0.361	0.374	-3.6#	99	0.00
46 T	Dibromomethane	0.277	0.278	-0.4	96	0.00
47 T	Bromodichloromethane	0.524	0.535	-2.1	96	0.00
48 T	Methyl methacrylate	0.410	0.414	-1.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032589.D
 Acq On : 04 Nov 2022 20:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	84	0.01
50 S	Toluene-d8	1.240	1.227	1.0	95	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.502	-0.4	94	0.00
52 CM	Toluene	0.917	0.946	-3.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.535	0.549	-2.6	93	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.594	-3.8	96	0.00
55 T	1,1,2-Trichloroethane	0.386	0.394	-2.1	99	0.00
56 T	Ethyl methacrylate	0.556	0.599	-7.7	97	0.00
57 T	1,3-Dichloropropane	0.624	0.644	-3.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.282	-5.2	96	0.00
59 T	2-Hexanone	0.371	0.375	-1.1	93	0.00
60 T	Dibromochloromethane	0.422	0.416	1.4	93	0.00
61 T	1,2-Dibromoethane	0.409	0.429	-4.9	99	0.00
62 S	4-Bromofluorobenzene	0.484	0.507	-4.8	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.366	0.360	1.6	95	0.00
65 PM	Chlorobenzene	1.120	1.133	-1.2	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.427	-1.2	97	0.00
67 C	Ethyl Benzene	2.003	2.058	-2.7#	98	0.00
68 T	m/p-Xylenes	0.778	0.795	-2.2	97	0.00
69 T	o-Xylene	0.756	0.791	-4.6	99	0.00
70 T	Styrene	1.257	1.319	-4.9	98	0.00
71 P	Bromoform	0.333	0.319	4.2	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.962	3.904	1.5	97	0.00
74 T	N-amyl acetate	1.497	1.500	-0.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.272	1.211	4.8	95	0.00
76 T	1,2,3-Trichloropropane	1.112	1.077	3.1	93	0.00
77 T	Bromobenzene	0.971	0.961	1.0	100	0.00
78 T	n-propylbenzene	4.565	4.524	0.9	96	0.00
79 T	2-Chlorotoluene	2.734	2.705	1.1	97	0.00
80 T	1,3,5-Trimethylbenzene	3.386	3.383	0.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.350	0.313	10.6	84	0.00
82 T	4-Chlorotoluene	3.199	3.222	-0.7	98	0.00
83 T	tert-Butylbenzene	3.381	3.324	1.7	96	0.00
84 T	1,2,4-Trimethylbenzene	3.368	3.390	-0.7	96	0.00
85 T	sec-Butylbenzene	4.068	4.019	1.2	94	0.00
86 T	p-Isopropyltoluene	3.452	3.432	0.6	93	0.00
87 T	1,3-Dichlorobenzene	1.832	1.833	-0.1	99	0.00
88 T	1,4-Dichlorobenzene	1.901	1.850	2.7	97	0.00
89 T	n-Butylbenzene	3.003	2.996	0.2	93	0.00
90 T	Hexachloroethane	0.564	0.512	9.2	86	0.00
91 T	1,2-Dichlorobenzene	1.807	1.803	0.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.244	12.5	83	0.00
93 T	1,2,4-Trichlorobenzene	1.178	1.136	3.6	97	0.00
94 T	Hexachlorobutadiene	0.492	0.440	10.6	91	0.00
95 T	Naphthalene	3.925	3.785	3.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032589.D
Acq On : 04 Nov 2022 20:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.197	1.137	5.0	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6