

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX110422\
 Data File : VX032589.D
 Acq On : 04 Nov 2022 20:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	54.885	-9.8	96	0.00
3 P	Chloromethane	50.000	52.298	-4.6	95	0.00
4 C	Vinyl Chloride	50.000	54.002	-8.0#	96	0.00
5 T	Bromomethane	50.000	41.028	17.9	74	0.00
6 T	Chloroethane	50.000	52.543	-5.1	103	0.00
7 T	Trichlorofluoromethane	50.000	51.425	-2.8	95	0.00
8 T	Diethyl Ether	50.000	52.942	-5.9	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.229	-4.5	100	0.00
10 T	Methyl Iodide	50.000	42.173	15.7	78	0.00
11 T	Tert butyl alcohol	250.000	240.791	3.7	91	0.01
12 CM	1,1-Dichloroethene	50.000	51.604	-3.2#	97	0.00
13 T	Acrolein	250.000	289.715	-15.9	113	0.00
14 T	Allyl chloride	50.000	52.352	-4.7	97	0.00
15 T	Acrylonitrile	250.000	253.306	-1.3	99	0.00
16 T	Acetone	250.000	248.477	0.6	101	0.00
17 T	Carbon Disulfide	50.000	43.039	13.9	78	0.00
18 T	Methyl Acetate	50.000	55.079	-10.2	104	0.00
19 T	Methyl tert-butyl Ether	50.000	53.587	-7.2	103	0.00
20 T	Methylene Chloride	50.000	50.397	-0.8	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.225	-0.5	96	0.00
22 T	Diisopropyl ether	50.000	53.215	-6.4	103	0.00
23 T	Vinyl Acetate	250.000	255.302	-2.1	96	0.00
24 P	1,1-Dichloroethane	50.000	52.558	-5.1	100	0.00
25 T	2-Butanone	250.000	243.407	2.6	95	0.00
26 T	2,2-Dichloropropane	50.000	47.257	5.5	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.328	-4.7	98	0.00
28 T	Bromochloromethane	50.000	47.646	4.7	95	0.00
29 T	Tetrahydrofuran	250.000	243.129	2.7	93	0.00
30 C	Chloroform	50.000	52.724	-5.4#	101	0.00
31 T	Cyclohexane	50.000	50.311	-0.6	93	0.00
32 T	1,1,1-Trichloroethane	50.000	52.431	-4.9	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.482	-1.0	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	50.700	-1.4	98	0.00
36 T	1,1-Dichloropropene	50.000	51.333	-2.7	97	0.00
37 T	Ethyl Acetate	50.000	47.074	5.9	95	0.00
38 T	Carbon Tetrachloride	50.000	50.449	-0.9	94	0.00
39 T	Methylcyclohexane	50.000	46.829	6.3	85	0.00
40 TM	Benzene	50.000	51.121	-2.2	97	0.00
41 T	Methacrylonitrile	50.000	54.073	-8.1	103	0.00
42 TM	1,2-Dichloroethane	50.000	52.433	-4.9	99	0.00
43 T	Isopropyl Acetate	50.000	50.564	-1.1	94	0.00
44 TM	Trichloroethene	50.000	50.299	-0.6	97	0.00
45 C	1,2-Dichloropropane	50.000	51.818	-3.6#	99	0.00
46 T	Dibromomethane	50.000	50.157	-0.3	96	0.00
47 T	Bromodichloromethane	50.000	50.976	-2.0	96	0.00
48 T	Methyl methacrylate	50.000	50.478	-1.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
 Data File : VX032589.D
 Acq On : 04 Nov 2022 20:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	905.391	9.5	84	0.01
50 S	Toluene-d8	50.000	49.459	1.1	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.163	-0.5	94	0.00
52 CM	Toluene	50.000	51.571	-3.1#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	51.258	-2.5	93	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.927	-3.9	96	0.00
55 T	1,1,2-Trichloroethane	50.000	51.077	-2.2	99	0.00
56 T	Ethyl methacrylate	50.000	53.879	-7.8	97	0.00
57 T	1,3-Dichloropropane	50.000	51.647	-3.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.856	-5.5	96	0.00
59 T	2-Hexanone	250.000	252.423	-1.0	93	0.00
60 T	Dibromochloromethane	50.000	49.305	1.4	93	0.00
61 T	1,2-Dibromoethane	50.000	52.365	-4.7	99	0.00
62 S	4-Bromofluorobenzene	50.000	52.353	-4.7	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	49.209	1.6	95	0.00
65 PM	Chlorobenzene	50.000	50.588	-1.2	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.703	-1.4	97	0.00
67 C	Ethyl Benzene	50.000	51.372	-2.7#	98	0.00
68 T	m/p-Xylenes	100.000	102.166	-2.2	97	0.00
69 T	o-Xylene	50.000	52.300	-4.6	99	0.00
70 T	Styrene	50.000	52.461	-4.9	98	0.00
71 P	Bromoform	50.000	47.832	4.3	88	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	49.269	1.5	97	0.00
74 T	N-ethyl acetate	50.000	50.105	-0.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.620	4.8	95	0.00
76 T	1,2,3-Trichloropropane	50.000	48.412	3.2	93	0.00
77 T	Bromobenzene	50.000	49.476	1.0	100	0.00
78 T	n-propylbenzene	50.000	49.556	0.9	96	0.00
79 T	2-Chlorotoluene	50.000	49.464	1.1	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.952	0.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.739	10.5	84	0.00
82 T	4-Chlorotoluene	50.000	50.356	-0.7	98	0.00
83 T	tert-Butylbenzene	50.000	49.168	1.7	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.323	-0.6	96	0.00
85 T	sec-Butylbenzene	50.000	49.393	1.2	94	0.00
86 T	p-Isopropyltoluene	50.000	49.714	0.6	93	0.00
87 T	1,3-Dichlorobenzene	50.000	50.028	-0.1	99	0.00
88 T	1,4-Dichlorobenzene	50.000	48.675	2.7	97	0.00
89 T	n-Butylbenzene	50.000	49.880	0.2	93	0.00
90 T	Hexachloroethane	50.000	45.371	9.3	86	0.00
91 T	1,2-Dichlorobenzene	50.000	49.892	0.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	43.754	12.5	83	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.222	3.6	97	0.00
94 T	Hexachlorobutadiene	50.000	44.750	10.5	91	0.00
95 T	Naphthalene	50.000	48.214	3.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110422\
Data File : VX032589.D
Acq On : 04 Nov 2022 20:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 05 00:40:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.462	5.1	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6