

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
 Data File : VX043679.D
 Acq On : 04 Nov 2024 13:04
 Operator : JC/MD
 Sample : VIBLK671
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK671

Quant Time: Nov 05 03:37:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 05 03:28:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	371559	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	311611	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	135983	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	63251	31.954	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	63.900%		
7) Chloroethane-d5	1.642	69	32062	37.683	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	75.360%		
11) 1,1-Dichloroethene-d2	2.294	65	31685	34.675	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	69.340%		
21) 2-Butanone-d5	4.459	46	132095	101.922	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	101.920%		
24) Chloroform-d	5.050	84	177981	41.709	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.420%		
26) 1,2-Dichloroethane-d4	5.952	65	119570	44.109	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.220%		
32) Benzene-d6	5.964	84	343334	42.616	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.240%		
36) 1,2-Dichloropropane-d6	7.306	67	111247	45.387	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.780%		
41) Toluene-d8	8.647	98	304970	41.671	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.340%		
43) trans-1,3-Dichloroprop...	8.952	79	52271	44.593	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	89.180%		
47) 2-Hexanone-d5	9.385	63	99780	102.865	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	102.860%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	157461	50.069	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	100.140%		
66) 1,2-Dichlorobenzene-d4	12.317	152	113704	48.260	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.520%		
Target Compounds					Qvalue	
14) Carbon disulfide	2.496	76	8299	1.894	ug/L	98
70) 1,2,4-trichlorobenzene	13.591	180	3687	1.454	ug/L	98
72) 1,2,3-Trichlorobenzene	13.963	180	3986	1.521	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
Data File : VX043679.D
Acq On : 04 Nov 2024 13:04
Operator : JC/MD
Sample : VIBLK671
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK671

Quant Time: Nov 05 03:37:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 05 03:28:26 2024
Response via : Initial Calibration

