

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
 Data File : VX043682.D
 Acq On : 04 Nov 2024 15:09
 Operator : JC/MD
 Sample : VX1104WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK716

Quant Time: Nov 05 03:46:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 05 03:44:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	345507	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	297007	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	128983	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	79674	43.286	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.580%
7) Chloroethane-d5	1.642	69	35048	44.298	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.294	65	36891	43.416	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.840%
21) 2-Butanone-d5	4.465	46	112031	92.959	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.960%
24) Chloroform-d	5.050	84	165757	41.774	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.540%
26) 1,2-Dichloroethane-d4	5.952	65	109708	43.522	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.040%
32) Benzene-d6	5.964	84	345151	44.948	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.900%
36) 1,2-Dichloropropane-d6	7.306	67	104541	44.749	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.500%
41) Toluene-d8	8.647	98	304925	43.713	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.420%
43) trans-1,3-Dichloroprop...	8.952	79	47797	42.781	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.560%
47) 2-Hexanone-d5	9.384	63	90117	97.471	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.470%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	134790	44.968	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.940%
66) 1,2-Dichlorobenzene-d4	12.317	152	105041	47.003	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
Data File : VX043682.D
Acq On : 04 Nov 2024 15:09
Operator : JC/MD
Sample : VX1104WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK716

Quant Time: Nov 05 03:46:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 05 03:44:12 2024
Response via : Initial Calibration

