

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
 Data File : VX043684.D
 Acq On : 04 Nov 2024 15:55
 Operator : JC/MD
 Sample : P4621-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BH6M7

Quant Time: Nov 05 03:46:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 05 03:44:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	337456	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	293634	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	129438	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	70168	39.031	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.060%
7) Chloroethane-d5	1.642	69	35283	45.659	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.320%
11) 1,1-Dichloroethene-d2	2.294	65	34536	41.614	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.220%
21) 2-Butanone-d5	4.465	46	144243	122.543	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.540%
24) Chloroform-d	5.056	84	190279	49.098	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.952	65	124841	50.707	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.420%
32) Benzene-d6	5.971	84	366722	48.306	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.620%
36) 1,2-Dichloropropane-d6	7.306	67	121341	52.537	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.080%
41) Toluene-d8	8.647	98	319278	46.297	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.600%
43) trans-1,3-Dichloroprop...	8.952	79	53605	48.531	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.060%
47) 2-Hexanone-d5	9.385	63	110238	120.604	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.600%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	171799	57.973	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	115.940%
66) 1,2-Dichlorobenzene-d4	12.317	152	124702	55.604	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110424\
Data File : VX043684.D
Acq On : 04 Nov 2024 15:55
Operator : JC/MD
Sample : P4621-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BH6M7

Quant Time: Nov 05 03:46:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 05 03:44:12 2024
Response via : Initial Calibration

