

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038923.D
 Acq On : 18 Nov 2023 13:12
 Operator : JC/MD
 Sample : 05404-21DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20231109DL

Quant Time: Nov 20 04:33:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	102125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202031	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78396	47.640	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.280%
35) Dibromofluoromethane	5.385	113	65614	44.968	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.940%
50) Toluene-d8	8.647	98	257829	46.184	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.360%
62) 4-Bromofluorobenzene	11.079	95	107295	48.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.020%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	35610	28.245	ug/l	Qvalue 97
20) Methylene Chloride	2.788	84	5389	3.413	ug/l #	96
44) Trichloroethene	7.129	130	34610	22.988	ug/l	89
64) Tetrachloroethene	9.275	164	2974	2.371	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038923.D
Acq On : 18 Nov 2023 13:12
Operator : JC/MD
Sample : 05404-21DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20231109DL

Quant Time: Nov 20 04:33:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

