

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038924.D
 Acq On : 18 Nov 2023 13:36
 Operator : JC/MD
 Sample : 05404-22DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20231109DL

Quant Time: Nov 20 04:34:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84021	49.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.980%
35) Dibromofluoromethane	5.385	113	68963	44.818	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.640%
50) Toluene-d8	8.646	98	269315	45.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.500%#
62) 4-Bromofluorobenzene	11.079	95	113422	48.128	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.260%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.324	101	1213	0.942	ug/l	88
20) Methylene Chloride	2.788	84	4737	2.937	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	1088	0.651	ug/l	77
44) Trichloroethene	7.128	130	55523	34.971	ug/l	100
64) Tetrachloroethene	9.274	164	2793	2.093	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038924.D
Acq On : 18 Nov 2023 13:36
Operator : JC/MD
Sample : 05404-22DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20231109DL

Quant Time: Nov 20 04:34:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

