

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038932.D
 Acq On : 18 Nov 2023 16:41
 Operator : JC/MD
 Sample : 05404-31
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20231110

Quant Time: Nov 20 04:37:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199429	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78890	49.723	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.440%
35) Dibromofluoromethane	5.385	113	64572	44.832	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.660%
50) Toluene-d8	8.647	98	251042	45.555	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.100%#
62) 4-Bromofluorobenzene	11.079	95	103132	46.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	4128	3.396	ug/l	97
12) 1,1-Dichloroethene	2.313	96	453	0.369	ug/l #	74
30) Chloroform	5.105	83	1161	0.472	ug/l #	81
44) Trichloroethene	7.129	130	532695	358.437	ug/l	96
64) Tetrachloroethene	9.275	164	1352	1.085	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038932.D
Acq On : 18 Nov 2023 16:41
Operator : JC/MD
Sample : 05404-31
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20231110

Quant Time: Nov 20 04:37:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

