

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038934.D
 Acq On : 18 Nov 2023 17:27
 Operator : JC/MD
 Sample : 05404-30
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20231110

Quant Time: Nov 20 04:37:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	98939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79622	49.943	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.880%
35) Dibromofluoromethane	5.385	113	65102	44.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.020%
50) Toluene-d8	8.647	98	253602	45.318	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.640%#
62) 4-Bromofluorobenzene	11.079	95	107179	47.844	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.680%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	2775	2.272	ug/l	90
12) 1,1-Dichloroethene	2.325	96	809	0.656	ug/l #	76
27) cis-1,2-Dichloroethene	4.501	96	2592	1.634	ug/l	87
44) Trichloroethene	7.123	130	423213	280.425	ug/l	94
64) Tetrachloroethene	9.275	164	2745	2.178	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038934.D
Acq On : 18 Nov 2023 17:27
Operator : JC/MD
Sample : 05404-30
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20231110

Quant Time: Nov 20 04:37:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

