

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038937.D
 Acq On : 18 Nov 2023 18:37
 Operator : JC/MD
 Sample : 05428-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20231113

Quant Time: Nov 20 04:38:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107621	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81749	49.024	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.040%
35) Dibromofluoromethane	5.385	113	67535	45.315	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	8.647	98	264513	46.388	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.780%
62) 4-Bromofluorobenzene	11.079	95	114616	50.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038937.D
Acq On : 18 Nov 2023 18:37
Operator : JC/MD
Sample : 05428-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20231113

Quant Time: Nov 20 04:38:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

