

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038941.D
 Acq On : 18 Nov 2023 20:09
 Operator : JC/MD
 Sample : 05404-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20231109

Quant Time: Nov 20 04:40:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212390	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83560	50.116	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.240%
35) Dibromofluoromethane	5.385	113	67549	44.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.080%
50) Toluene-d8	8.647	98	269759	45.964	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.920%#
62) 4-Bromofluorobenzene	11.079	95	112868	48.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	884	0.692	ug/l	# 93
20) Methylene Chloride	2.788	84	1020	0.637	ug/l	84
44) Trichloroethene	7.123	130	51844	32.756	ug/l	94
64) Tetrachloroethene	9.275	164	3032	2.284	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038941.D
Acq On : 18 Nov 2023 20:09
Operator : JC/MD
Sample : 05404-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20231109

Quant Time: Nov 20 04:40:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

