

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038918.D
 Acq On : 18 Nov 2023 11:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	66	0.00
2 T	Dichlorodifluoromethane	0.583	0.614	-5.3	66	0.00
3 P	Chloromethane	0.761	0.664	12.7	60	0.00
4 C	Vinyl Chloride	0.734	0.685	6.7#	63	0.00
5 T	Bromomethane	0.303	0.349	-15.2	67	0.00
6 T	Chloroethane	0.419	0.425	-1.4	67	0.00
7 T	Trichlorofluoromethane	0.931	0.962	-3.3	69	0.00
8 T	Diethyl Ether	0.405	0.405	0.0	70	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.582	5.7	64	0.00
10 T	Methyl Iodide	0.751	0.716	4.7	62	0.00
11 T	Tert butyl alcohol	0.224	0.237	-5.8	72	0.00
12 CM	1,1-Dichloroethene	0.623	0.571	8.3#	62	0.00
13 T	Acrolein	0.119	0.135	-13.4	82	0.00
14 T	Allyl chloride	1.248	1.173	6.0	64	0.00
15 T	Acrylonitrile	0.473	0.473	0.0	66	0.00
16 T	Acetone	0.429	0.516	-20.3	82	0.00
17 T	Carbon Disulfide	1.930	1.570	18.7	58	0.00
18 T	Methyl Acetate	2.037	2.094	-2.8	68	0.00
19 T	Methyl tert-butyl Ether	2.192	2.228	-1.6	68	0.00
20 T	Methylene Chloride	0.773	0.710	8.2	67	0.00
21 T	trans-1,2-Dichloroethene	0.688	0.635	7.7	63	0.00
22 T	Diisopropyl ether	2.289	2.265	1.0	66	0.00
23 T	Vinyl Acetate	1.649	1.656	-0.4	66	0.00
24 P	1,1-Dichloroethane	1.280	1.225	4.3	65	0.00
25 T	2-Butanone	0.696	0.732	-5.2	71	0.00
26 T	2,2-Dichloropropane	1.110	1.057	4.8	64	0.00
27 T	cis-1,2-Dichloroethene	0.802	0.754	6.0	65	0.00
28 T	Bromochloromethane	0.615	0.606	1.5	67	0.00
29 T	Tetrahydrofuran	0.442	0.454	-2.7	67	0.00
30 C	Chloroform	1.250	1.223	2.2#	67	0.00
31 T	Cyclohexane	1.150	1.077	6.3	63	0.00
32 T	1,1,1-Trichloroethane	1.083	1.063	1.8	67	0.00
33 S	1,2-Dichloroethane-d4	0.806	0.771	4.3	64	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
35 S	Dibromofluoromethane	0.361	0.342	5.3	61	0.00
36 T	1,1-Dichloropropene	0.518	0.496	4.2	63	0.00
37 T	Ethyl Acetate	0.674	0.702	-4.2	67	0.00
38 T	Carbon Tetrachloride	0.478	0.496	-3.8	68	0.00
39 T	Methylcyclohexane	0.675	0.632	6.4	63	0.00
40 TM	Benzene	1.531	1.487	2.9	64	0.00
41 T	Methacrylonitrile	0.350	0.364	-4.0	65	0.00
42 TM	1,2-Dichloroethane	0.507	0.534	-5.3	69	0.00
43 T	Isopropyl Acetate	1.078	1.121	-4.0	67	0.00
44 TM	Trichloroethene	0.373	0.372	0.3	65	0.00
45 C	1,2-Dichloropropane	0.412	0.413	-0.2#	64	0.00
46 T	Dibromomethane	0.275	0.283	-2.9	67	0.00
47 T	Bromodichloromethane	0.545	0.555	-1.8	68	0.00
48 T	Methyl methacrylate	0.523	0.520	0.6	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
 Data File : VX038918.D
 Acq On : 18 Nov 2023 11:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.012	-20.0	76	0.00
50 S	Toluene-d8	1.382	1.288	6.8	60	0.00
51 T	4-Methyl-2-Pentanone	0.689	0.726	-5.4	68	0.00
52 CM	Toluene	0.950	0.939	1.2#	65	0.00
53 T	t-1,3-Dichloropropene	0.648	0.649	-0.2	65	0.00
54 T	cis-1,3-Dichloropropene	0.669	0.676	-1.0	65	0.00
55 T	1,1,2-Trichloroethane	0.396	0.399	-0.8	66	0.00
56 T	Ethyl methacrylate	0.671	0.694	-3.4	66	0.00
57 T	1,3-Dichloropropane	0.656	0.670	-2.1	67	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.317	-1.6	78	0.00
59 T	2-Hexanone	0.552	0.597	-8.2	69	0.00
60 T	Dibromochloromethane	0.403	0.422	-4.7	69	0.00
61 T	1,2-Dibromoethane	0.416	0.426	-2.4	67	0.00
62 S	4-Bromofluorobenzene	0.553	0.506	8.5	60	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.311	0.319	-2.6	69	0.00
65 PM	Chlorobenzene	1.077	1.072	0.5	66	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.393	-1.6	68	0.00
67 C	Ethyl Benzene	1.979	1.988	-0.5#	66	0.00
68 T	m/p-Xylenes	0.749	0.749	0.0	66	0.00
69 T	o-Xylene	0.741	0.738	0.4	66	0.00
70 T	Styrene	1.251	1.244	0.6	66	0.00
71 P	Bromoform	0.323	0.345	-6.8	69	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
73 T	Isopropylbenzene	3.896	3.814	2.1	67	0.00
74 T	N-amyl acetate	2.059	2.069	-0.5	66	0.00
75 P	1,1,2,2-Tetrachloroethane	1.521	1.492	1.9	67	0.00
76 T	1,2,3-Trichloropropane	1.247	1.254	-0.6	67	0.00
77 T	Bromobenzene	0.875	0.856	2.2	67	0.00
78 T	n-propylbenzene	4.795	4.719	1.6	66	0.00
79 T	2-Chlorotoluene	2.846	2.748	3.4	67	0.00
80 T	1,3,5-Trimethylbenzene	3.276	3.291	-0.5	67	0.00
81 T	trans-1,4-Dichloro-2-butene	0.562	0.536	4.6	63	0.00
82 T	4-Chlorotoluene	3.281	3.208	2.2	66	0.00
83 T	tert-Butylbenzene	3.226	3.203	0.7	68	0.00
84 T	1,2,4-Trimethylbenzene	3.315	3.286	0.9	67	0.00
85 T	sec-Butylbenzene	4.220	4.182	0.9	67	0.00
86 T	p-Isopropyltoluene	3.387	3.396	-0.3	68	0.00
87 T	1,3-Dichlorobenzene	1.714	1.697	1.0	67	0.00
88 T	1,4-Dichlorobenzene	1.754	1.725	1.7	68	0.00
89 T	n-Butylbenzene	3.418	3.377	1.2	66	0.00
90 T	Hexachloroethane	0.676	0.662	2.1	66	0.00
91 T	1,2-Dichlorobenzene	1.626	1.631	-0.3	69	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.373	0.387	-3.8	69	0.00
93 T	1,2,4-Trichlorobenzene	1.089	1.093	-0.4	68	0.00
94 T	Hexachlorobutadiene	0.386	0.384	0.5	66	0.00
95 T	Naphthalene	4.058	4.016	1.0	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038918.D
Acq On : 18 Nov 2023 11:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.033	1.049	-1.5	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6