

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX111823\
 Data File : VX038918.D
 Acq On : 18 Nov 2023 11:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	66	0.00
2 T	Dichlorodifluoromethane	50.000	52.599	-5.2	66	0.00
3 P	Chloromethane	50.000	43.588	12.8	60	0.00
4 C	Vinyl Chloride	50.000	46.669	6.7#	63	0.00
5 T	Bromomethane	50.000	57.569	-15.1	67	0.00
6 T	Chloroethane	50.000	50.710	-1.4	67	0.00
7 T	Trichlorofluoromethane	50.000	51.687	-3.4	69	0.00
8 T	Diethyl Ether	50.000	49.943	0.1	70	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.162	5.7	64	0.00
10 T	Methyl Iodide	50.000	47.647	4.7	62	0.00
11 T	Tert butyl alcohol	250.000	264.114	-5.6	72	0.00
12 CM	1,1-Dichloroethene	50.000	45.820	8.4#	62	0.00
13 T	Acrolein	250.000	283.783	-13.5	82	0.00
14 T	Allyl chloride	50.000	47.018	6.0	64	0.00
15 T	Acrylonitrile	250.000	249.753	0.1	66	0.00
16 T	Acetone	250.000	300.666	-20.3	82	0.00
17 T	Carbon Disulfide	50.000	40.673	18.7	58	0.00
18 T	Methyl Acetate	50.000	51.386	-2.8	68	0.00
19 T	Methyl tert-butyl Ether	50.000	50.831	-1.7	68	0.00
20 T	Methylene Chloride	50.000	45.924	8.2	67	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.179	7.6	63	0.00
22 T	Diisopropyl ether	50.000	49.470	1.1	66	0.00
23 T	Vinyl Acetate	250.000	251.114	-0.4	66	0.00
24 P	1,1-Dichloroethane	50.000	47.859	4.3	65	0.00
25 T	2-Butanone	250.000	263.188	-5.3	71	0.00
26 T	2,2-Dichloropropane	50.000	47.584	4.8	64	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.037	5.9	65	0.00
28 T	Bromochloromethane	50.000	49.293	1.4	67	0.00
29 T	Tetrahydrofuran	250.000	257.184	-2.9	67	0.00
30 C	Chloroform	50.000	48.906	2.2#	67	0.00
31 T	Cyclohexane	50.000	46.812	6.4	63	0.00
32 T	1,1,1-Trichloroethane	50.000	49.077	1.8	67	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.840	4.3	64	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	64	0.00
35 S	Dibromofluoromethane	50.000	47.365	5.3	61	0.00
36 T	1,1-Dichloropropene	50.000	47.961	4.1	63	0.00
37 T	Ethyl Acetate	50.000	52.063	-4.1	67	0.00
38 T	Carbon Tetrachloride	50.000	51.892	-3.8	68	0.00
39 T	Methylcyclohexane	50.000	46.822	6.4	63	0.00
40 TM	Benzene	50.000	48.562	2.9	64	0.00
41 T	Methacrylonitrile	50.000	52.051	-4.1	65	0.00
42 TM	1,2-Dichloroethane	50.000	52.690	-5.4	69	0.00
43 T	Isopropyl Acetate	50.000	52.007	-4.0	67	0.00
44 TM	Trichloroethene	50.000	49.876	0.2	65	0.00
45 C	1,2-Dichloropropane	50.000	50.154	-0.3#	64	0.00
46 T	Dibromomethane	50.000	51.536	-3.1	67	0.00
47 T	Bromodichloromethane	50.000	50.884	-1.8	68	0.00
48 T	Methyl methacrylate	50.000	49.739	0.5	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX111823\
 Data File : VX038918.D
 Acq On : 18 Nov 2023 11:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1141.850	-14.2	76	0.00
50 S	Toluene-d8	50.000	46.618	6.8	60	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.727	-5.5	68	0.00
52 CM	Toluene	50.000	49.403	1.2#	65	0.00
53 T	t-1,3-Dichloropropene	50.000	50.050	-0.1	65	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.507	-1.0	65	0.00
55 T	1,1,2-Trichloroethane	50.000	50.372	-0.7	66	0.00
56 T	Ethyl methacrylate	50.000	51.775	-3.5	66	0.00
57 T	1,3-Dichloropropane	50.000	51.068	-2.1	67	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.652	-1.5	78	0.00
59 T	2-Hexanone	250.000	270.516	-8.2	69	0.00
60 T	Dibromochloromethane	50.000	52.359	-4.7	69	0.00
61 T	1,2-Dibromoethane	50.000	51.152	-2.3	67	0.00
62 S	4-Bromofluorobenzene	50.000	45.789	8.4	60	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	64	0.00
64 T	Tetrachloroethene	50.000	51.300	-2.6	69	0.00
65 PM	Chlorobenzene	50.000	49.751	0.5	66	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.663	-1.3	68	0.00
67 C	Ethyl Benzene	50.000	50.224	-0.4#	66	0.00
68 T	m/p-Xylenes	100.000	100.018	-0.0	66	0.00
69 T	o-Xylene	50.000	49.798	0.4	66	0.00
70 T	Styrene	50.000	49.729	0.5	66	0.00
71 P	Bromoform	50.000	53.419	-6.8	69	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	64	0.00
73 T	Isopropylbenzene	50.000	48.947	2.1	67	0.00
74 T	N-ethyl acetate	50.000	50.226	-0.5	66	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.062	1.9	67	0.00
76 T	1,2,3-Trichloropropane	50.000	50.278	-0.6	67	0.00
77 T	Bromobenzene	50.000	48.889	2.2	67	0.00
78 T	n-propylbenzene	50.000	49.205	1.6	66	0.00
79 T	2-Chlorotoluene	50.000	48.282	3.4	67	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.228	-0.5	67	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.607	4.8	63	0.00
82 T	4-Chlorotoluene	50.000	48.876	2.2	66	0.00
83 T	tert-Butylbenzene	50.000	49.650	0.7	68	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.556	0.9	67	0.00
85 T	sec-Butylbenzene	50.000	49.549	0.9	67	0.00
86 T	p-Isopropyltoluene	50.000	50.132	-0.3	68	0.00
87 T	1,3-Dichlorobenzene	50.000	49.486	1.0	67	0.00
88 T	1,4-Dichlorobenzene	50.000	49.165	1.7	68	0.00
89 T	n-Butylbenzene	50.000	49.401	1.2	66	0.00
90 T	Hexachloroethane	50.000	48.946	2.1	66	0.00
91 T	1,2-Dichlorobenzene	50.000	50.161	-0.3	69	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.946	-3.9	69	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.182	-0.4	68	0.00
94 T	Hexachlorobutadiene	50.000	49.725	0.5	66	0.00
95 T	Naphthalene	50.000	49.485	1.0	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111823\
Data File : VX038918.D
Acq On : 18 Nov 2023 11:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 04:30:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.779	-1.6	68	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6