

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112919\
 Data File : VX013693.D
 Acq On : 29 Nov 2019 11:55
 Operator : JC/SP
 Sample : VX1129WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1129WBL01

Quant Time: Nov 29 22:50:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	703871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1092680	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	964905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	431617	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	394703	48.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.62%	
35) Dibromofluoromethane	5.48	113	319638	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	
50) Toluene-d8	8.71	98	1291073	49.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.12%	
62) 4-Bromofluorobenzene	11.13	95	429090	45.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112919\
Data File : VX013693.D
Acq On : 29 Nov 2019 11:55
Operator : JC/SP
Sample : VX1129WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1129WBL01

Quant Time: Nov 29 22:50:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

