

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112919\
 Data File : VX013706.D
 Acq On : 29 Nov 2019 17:10
 Operator : JC/SP
 Sample : K5965-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S39-0194

Quant Time: Dec 02 06:56:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	684522	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1057611	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	934887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	411720	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	379641	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
35) Dibromofluoromethane	5.49	113	314234	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
50) Toluene-d8	8.71	98	1246031	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.14	95	413633	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	28813	3.535	ug/l	88
32) 1,1,1-Trichloroethane	5.48	97	8908	0.845	ug/l #	48
44) Trichloroethene	7.21	130	7343	0.915	ug/l	86
64) Tetrachloroethene	9.34	164	4308	0.618	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112919\
Data File : VX013706.D
Acq On : 29 Nov 2019 17:10
Operator : JC/SP
Sample : K5965-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S39-0194

Quant Time: Dec 02 06:56:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

